



EA MLA Signatory
Český institut pro akreditaci, o.p.s.
Hájkova 2747/22, Žižkov, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

CERTIFICATE OF ACCREDITATION

No. 42/2025

VENTURI Calibration Services spol. s r.o.
with registered office Blatnice pod Svatým Antonínkem 866, 696 71,
Company Registration No. 26270854

for the Calibration Laboratory No. 2366
VENTURI Calibration Services spol. s r.o.

Scope of accreditation:

Calibration of flow meters to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Conformity Assessment Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Accredited Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 89/2020 of 13/02/2020, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **29/01/2030**

Prague: 29/01/2025



Jan Velíšek
Director of the Department
of Testing and Calibration Laboratories
Czech Accreditation Institute



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

VENTURI Calibration Services, spol. s r.o.
CAB number 2366, VENTURI Calibration Services spol. s r.o.
Blatnice pod Svatým Antonínkem 866, PSČ 696 71

CMC for the field of measured quantity: Flow

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.					
1	Flow meters, volume meters and water meters	1.0 m ³ /h 40 m ³ /h 11.1 kg/s	to	150 m ³ /h 4,245 m ³ /h 1,166 kg/s	20 °C to 80 °C 35 °C to 80 °C 35 °C to 80 °C	0.20 % 0.20 % 0.10 %	Volumetric method Volumetric method Gravimetric method	KP 03 KP 01 KP 01	
2	Metering runs with flow restriction elements	1.0 m ³ /h 40 m ³ /h 11.1 kg/s	to	150 m ³ /h 4,245 m ³ /h 1,166 kg/s	20 °C to 80 °C 35 °C to 80 °C 35 °C to 80 °C	0.21 % 0.21 % 0.14 %	Volumetric method Volumetric method Gravimetric method	KP 04 KP 02 KP 02	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

